U.S. Department of Commerce

National Institute of Standards and Technology Gaithersburg, MD 20899

Certificate Number: 00-010

Page 1 of 2

National Type Evaluation Program

Certificate of Conformance

for Weighing and Measuring Devices

For:

Bench Scale or Weight Classifier

Digital Electronic Model: PS90 n_{max}: 3000

Capacity: 150 lb x 0.05 lb (60 kg x 0.02 kg) Platform: 15.75" x 19.75" (400 x 500 mm)

Accuracy Class: III

Submitted by:

Mettler-Toledo, Inc. P.O. Box 1705

Columbus, OH 43216 Tel: (614) 438-4393 Fax: (614) 438-4355 Contact: Darrell Flocken

Standard Features and Options

Semi-automatic (push-button) zero
Automatic zero setting mechanism (AZSM)
Initial zero setting mechanism (IZSM)
Remote printer capability
AC/DC adapter
External lb/kg conversion
RS-232 communication port
Stainless steel platter
Liquid crystal display (LCD)

Options: Ball top platter

Roller top platter

Load cells used: Mettler-Toledo MT1241-100, 100 kg capacity

Temperature Range: -10 °C to 40 °C (14 °F to 104 °F)

This device was evaluated under the National Type Evaluation Program (NTEP) and was found to comply with the applicable technical requirements of Handbook 44, "Specifications, Tolerances, and Other Technical Requirements for Weighing and Measuring Devices." Evaluation results and device characteristics necessary for inspection and use in commerce are on the following pages.

Effective Date: January 3, 2000

Gilbert M. Ugiansky, Ph.D. Chief, Office of Weights and Measures Issue Date: January 12, 2000

Note: The National Institute of Standards and Technology does not "approve", "recommend", or "endorse" any proprietary product or material, either as a single item or as a class or group. Results shall not be used in advertising or sales promotion to indicate explicit or implicit endorsement of the product by the Institute. (See NTEP Policy and Procedures.)

Certificate Number: 00-010

Page 2 of 2

Mettler-Toledo, Inc. Bench Scale or Weight Classifier Model: PS90

Application: General purpose bench and counter scale.

<u>Identification:</u> The required information is on an adhesive badge under the platter attached to the base of the scale.

<u>Sealing:</u> The scale can be sealed with a wire security seal threaded through a hole in the display mounting bracket and a drilled head screw that secures a plate covering the circuit board to prevent access to the calibration and configuration switch.

<u>Test Conditions:</u> The emphasis of the evaluation was on the device design, operation, marking requirements and compliance with influence factor requirements. Several increasing/decreasing load and shift tests were performed. The scale was tested over a temperature range of -10 °C to 40 °C (14 °F to 104 °F). A load of approximately one-half scale capacity was applied to the scale 100 184 times. The scale was tested periodically during this time. Tests were also conducted with a power supply of 100 VAC or 130 VAC.

The results of the evaluation indicate the device complies with applicable requirements of NIST Handbook 44

Type Evaluation Criteria Used: NIST Handbook 44, 2000 Edition

Tested By: E. Matthews (OH), W. West (OH)